

Certificate of Conformity

No.: LP23100029C02-01

The following products have been tested by us with the listed standards and found in compliance with the council LVD directive 2014/35/EU. It is possible to use CE marking to demonstrative for the compliance with this LVD Directive

Applicant : Shenzhen Green Digital Power Tech Co., LTD.
Address : #502, 5F BLDG 01, Daxin Xinmei Industrial Park, 21 Gongchang Road, Guangming District, Shenzhen, Guangdong, P.R. China
Product : Battery 4-in-1 Instrument Tester/Charger/Discharger/Maintainer
Trade Mark : N/A
Model No. : BM200-32, BM200-05
Rated Input : 100-240V~, 50/60Hz, 2.2A
Rated Output : 2-32V === 6.0A Max for model BM200-32;
2-5V === 30.0A Max for model BM200-05
Test Standards : EN 61010-1:2010+A1:2019 Safety requirements for electrical equipment for measurement, control, and laboratory use Part 1: General requirements



Manager: Haylen Li
Date: Dec. 18, 2023

Dongguan Lepont Testing Service Co., Ltd.

Tel.: 86-769-8308 6888 Fax.: 86-769-8309 6886
Room 102, Building 11, No. 7, Houjie Science And Technology Avenue, Houjie, Dongguan, Guangdong, China

The statement is based on a single evaluation of one sample of above mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab's logo.